

# F&P 950™ Respiratory Humidifier

## USER INSTRUCTIONS



## Contents

---

Indications for use .....	3
Operating principle .....	3
Package contents .....	4
F&P 950 Respiratory Humidifier setup .....	5
User interface.....	7
Alarms .....	10
Information and service menus.....	14
Cleaning and maintenance .....	17
Warnings, cautions and notes.....	17
Symbol definitions.....	19
Technical specifications .....	20
<i>Product specifications</i> .....	20
<i>Operating conditions</i> .....	21
<i>Storage conditions</i> .....	21

## Indications for use

---

The F&P 950 Respiratory Humidifier is intended to provide heat and humidity to respiratory gases delivered to patients.

## Operating principle

---

The F&P 950 Respiratory Humidifier provides heat and humidity to medical gases by passing the gas through a heated water chamber and heated breathing tubes.

The amount of heating is controlled based on the gas temperature measured at different parts of the humidifier.



## Package contents

---



**F&P 950 Heaterbase**  
(e.g. 950ANZ)



**Power cord**  
(e.g. 950XPI)


### Accessories to complete the F&P 950 Respiratory Humidifier



**F&P 950 Sensor Cartridge**  
(e.g. 950S02)



**Equipment mount**  
(e.g. 900MR030)




**F&P 950 Breathing Circuit Kit**  
(e.g. 950A81, 950N80)



**F&P 950 Expiratory Heater Wire Adapter**  
(e.g. 950X00)

## F&P 950 Respiratory Humidifier setup

Attach the power cord and power cord retainer to the heaterbase.



Attach the sensor cartridge to the heater base.



### WARNING


The heaterbase must be mounted on an equipment mount capable of supporting 4 kg. Failure to comply may result in damage to the equipment mount and heaterbase, and potentially cause serious patient harm.

### NOTES

- Ensure the heaterbase does not block access to the power supply outlet.
- Update the heaterbase software to Rev J (6.0.10) or later before attaching the 950S02 Sensor Cartridge.

## F&P 950 Respiratory Humidifier setup

Set up the breathing circuit as described in the user instructions provided with the selected breathing circuit kit.



When turning on the humidifier, an audible single beep sound should be heard.

## User interface

### Screen navigation



### Modes

The modes available will depend on the type of breathing circuit connected. The availability and operating principles for each mode are shown below.

#### Breathing Circuit Kit

#### Modes

Adult & Pediatric Breathing Circuit Kits

 **Invasive**

**Invasive mode** is intended for patients whose upper airways have been bypassed by either a tracheostomy or endotracheal tube.

 **Mask**

**Mask mode** is intended for patients whose upper airways have not been bypassed but are receiving gas via a face mask or similar.

 **Optiflow**

**Optiflow™ mode** is intended for patients who require respiratory therapy through an Optiflow interface.

Neonatal Breathing Circuit Kit (Additional modes disabled)

 **Neonatal**

**Neonatal mode** is intended for neonates who require respiratory support.

Neonatal Breathing Circuit Kit (Additional modes enabled)

 **Invasive**

**Invasive mode** is intended for patients whose upper airways have been bypassed by either a tracheostomy or endotracheal tube.

 **CPAP | NIV**

**CPAP | NIV mode** is intended for patients whose upper airways have not been bypassed and are receiving positive pressure therapy through a sealed or nasal interface.

 **Optiflow**

**Optiflow mode** is intended for patients who require respiratory therapy through an Optiflow interface.

Optiflow Oxygen Kit

 **Optiflow**

**Optiflow mode** is intended for patients who require respiratory therapy through an Optiflow interface.

## User interface

---

When multiple modes exist for a type of breathing circuit kit, selection can be accessed via the drop-down menu button.



## User interface

### Comfort settings

With an adult or pediatric inspiratory limb connected, it is possible to change the target temperature in Mask and Optiflow modes, to provide conditions which may encourage patient comfort.

When additional neonatal modes are enabled, changing the target temperature in CPAP | NIV and Optiflow modes is also possible.



The available comfort settings are:

Adult & Pediatric			
Mode	Default	Medium	Low
Invasive	37 °C	-	-
Mask	31 °C	29 °C	27 °C
Optiflow	37 °C	35 °C	33 °C

Neonatal			
Mode	Default	Medium	Low
Neonatal	37 °C	-	-
Invasive*	37 °C	-	-
CPAP   NIV*	37 °C	34 °C	31 °C
Optiflow*	37 °C	35 °C	33 °C

\* with additional modes enabled

The humidifier will reset to the default set-point if the mode is changed or the humidifier is turned off and back on. It is possible for service personnel to change the default set-point for Mask, CPAP | NIV and Optiflow modes in the service menu.

## Alarms

---

### Alarm signals

The F&P 950 Respiratory Humidifier has visual and audible alarms to warn about interruptions to treatment. These alarms are generated by an intelligent alarm system, which processes information from the sensors and target settings of the unit and compares this information to pre-programmed limits.

#### Alarm type

#### Text or animation tutorial

Demonstrates corrective action – see “alarm conditions” table on the next page.

#### Mute/unmute alarm

2 minutes.



## Alarms

---

### Alarm conditions

All possible alarm conditions are listed on the following pages, and all are classified as medium or low priorities.

As the F&P 950 Respiratory Humidifier does not include patient monitoring, these alarms are considered technical indicators of humidifier performance. It is possible to have multiple alarm conditions occur simultaneously; under these conditions the humidifier uses an internal ranking system to display the highest-ranked alarm.

Medium priority alarms have been designed to be detectable within one meter of the heaterbase, with the alarm signal being three beeps repeated every five seconds.

Low priority alarms have been designed to be detectable within one meter of the heaterbase, with the alarm signal being one beep repeated every five seconds.

### Checking alarm system functionality

**⚠ WARNING:** Do not remove breathing circuit when connected to a patient. Failure to comply may compromise safety, including serious patient harm.

To check alarm functionality, remove the heated breathing tube at any time while the humidifier is powered on **but not connected to a patient**. This action should activate the “Disconnection” visual and audible alarms. If either signal is absent, do not use the humidifier. Contact your servicing department for assistance.

In the event of an unexpected shutdown, the humidifier shall resume the operating mode and alarm settings (except algorithm-based alarms) prior to the reset if the interruption is less than or equal to 30 seconds.

## Alarms

### Alarm Priority: Medium

ALARM CONDITIONS	REQUIRED ACTION
<p><b>The Disconnection alarm</b> activates when the humidifier detects a disconnection of the inspiratory circuit.</p> <p>Delay: &lt; 10 seconds</p>	<p>Connect inspiratory circuit and fully insert the chamber for complete connection.</p>
<p><b>The No Water alarm</b> activates when the humidifier detects that the chamber is empty or almost empty of water.</p> <p>The time-to-alarm signal generation is dependent on operating mode set-point and flow rates as these determine the water evaporation rate.</p> <p>Delay: &lt; 60 minutes</p>	<p>Replace the empty water bag.</p>
<p><b>The Check Setup alarm</b> activates when the humidifier detects a repeated elevated temperature condition at the chamber outlet.</p> <p>The alarm threshold is 43 °C.</p> <p>The time-to-alarm signal generation is dependent on the flow rates.</p> <p>Delay: &gt; 5 minutes</p>	<p>Check the dryline and expiratory limb connect to the correct ports on the flow source.</p>
<p><b>The Low Temperature alarm</b> activates when the humidifier detects a low temperature condition at the patient end or chamber outlet for a continuous period of time.</p> <p>The alarm threshold is 2 °C below the set-point temperature.</p> <p>The time-to-alarm signal generation is dependent on the flow rates.</p> <p>Delay: &gt; 10 minutes</p>	<p>Check the humidifier is receiving flow within the range stated in this user instruction.</p> <p>Check the humidifier setup.</p>
<p><b>The High Temperature alarm</b> activates when the humidifier detects a high temperature condition at the patient end.</p> <p>The alarm threshold is a patient end temperature of &gt; 43 °C.</p> <p>Delay: &lt; 30 seconds</p>	<p>Check the humidifier is receiving flow within the range stated in this user instruction.</p> <p>Check connections to the flow source.</p> <p>Check the humidifier setup.</p>
<p><b>The Cartridge Disconnection alarm</b> activates when the humidifier detects that the sensor cartridge is not electrically connected.</p> <p>Delay: &lt; 10 seconds</p>	<p>Connect the sensor cartridge.</p>
<p><b>The Tube Fault alarm</b> activates when the humidifier detects a potential fault in the breathing circuit.</p> <p>Delay: &lt; 10 seconds</p>	<p>Replace the breathing circuit when safe to do so.</p>

## Alarms

### Alarm Priority: Medium

ALARM CONDITIONS	REQUIRED ACTION
<p><b>The Service Required alarm</b> activates when the humidifier detects a potential fault that requires the humidifier to be serviced.</p> <p>Delay: 10 seconds to 5 minutes</p>	<p>Turn off the humidifier as soon as appropriate, remove from service, and contact a technician.</p>
<p><b>The Caution Indicator LED light</b> illuminates when the humidifier detects that there is a potential fault with the humidifier and the screen is not operational.</p> <p>Delay: &lt; 10 seconds</p>	<p>Turn off the humidifier as soon as appropriate, remove from service, and contact a technician.</p>
<p><b>The Cartridge Authentication alarm</b> activates when the humidifier does not recognize the sensor cartridge.</p> <p>If this occurs, the user may choose to press "Accept" to acknowledge that the sensor cartridge is not Fisher Paykel Healthcare approved.</p> <p>A sensor cartridge authentication failure icon will appear at the bottom right of the display.</p>	<p>To remove the sensor cartridge authentication failure icon, contact technician to replace sensor cartridge as soon as appropriate.</p>
<p><b>The Cartridge Service Life alarm</b> activated when the humidifier detects the sensor cartridge has exceeded the recommended service life.</p> <p>The sensor cartridge should be replaced at the next opportunity that it is safe to do so (when not in use by a patient).</p> <p>Delay: 15,000 hours of use or 7 years from the date of manufacture, whichever is earlier.</p> <p>If the alarm is paused, it will reappear 4 hours later.</p>	<p>Press "Pause Alarm" button to dismiss the alarm screen.</p> <p>Contact technician to replace sensor cartridge as soon as appropriate.</p>

### Alarm Priority: Low

ALARM CONDITIONS	REQUIRED ACTION
<p><b>The Check Adapter alarm</b> activates when the humidifier detects the expiratory heater wire adapter is disconnected.</p> <p><i>Note: By default this alarm is disabled. It can be enabled through the service menu.</i></p> <p>Delay: &lt; 20 seconds</p>	<p>Connect the expiratory heater wire adapter between the sensor cartridge and the expiratory circuit.</p>

## Alarms

### Information signals

Notification type



Usage

Notification content

Demonstrates corrective action – see “information signals” table below.

• Replace sensor cartridge. The cartridge will expire in 2 days.

Mute/unmute alarm


Remind me later

INFORMATION SIGNALS	POSSIBLE ACTIONS
<p><b>The Cartridge Service Life warning</b> activates when the humidifier detects the sensor cartridge is approaching the end of its recommended service life.</p> <p>At this point the sensor cartridge has one month of service life remaining and a sensor cartridge should be made available for replacement.</p> <p>Delay: 30 days prior to expiry and will reappear every 24 hours, or every 8 hours if less than 7 days remaining</p>	<p>Press “Remind me later” button to dismiss the warning screen.</p> <p>Contact technician to replace sensor cartridge as soon as appropriate.</p>

## Information and service menus

### Options screen

The "Options" screen contains additional information about the humidifier and can be accessed by pressing the "Menu" button. Tapping on each option enables navigation through the screens.




The servicing functions are password protected and should only be accessed by technical personnel. Refer to the Product Technical Manual for more information.

**NOTE:** The readings displayed in the Operating Functions page under the Information directory are additional information for troubleshooting purposes only. These values are not intended to be used to specify patient treatment or for patient diagnosis.

## Information and service menus

### Lock screen function

The F&P 950 Heaterbase screen can be locked to avoid unintentional changes to modes or settings. Follow the instruction below to enable or disable the feature:

STEP	INSTRUCTION	SCREENSHOT
1	Navigate to the "Options" screen by touching the menu icon in the bottom left corner of the "Main" screen.	
2	Press and hold the lock icon.  Hold down the icon until the countdown animation completes one full revolution.	  
3	When the screen is locked, a "lock" icon is displayed.	

4

To unlock the screen, tap the lock icon once.



The icon will change to "unlock".  
Press and hold the "unlock" icon.



Hold down the icon until the countdown animation completes one full revolution.



5

When unlocked, the humidifier will return to the main screen and the user will be able to change the mode or settings.



**NOTE:** Please refer queries relating to setup, troubleshooting, service, repair and unexpected operation of the humidifier or accessories, to your healthcare provider or local Fisher & Paykel Healthcare representative.

## Cleaning and maintenance

---

### Cleaning

Clean the heaterbase, sensor cartridge, or expiratory heater wire adapter using a cloth dampened with either isopropyl alcohol or neutral detergent. Always disconnect the humidifier from the power supply before cleaning.

#### NOTES:

- Do not immerse or autoclave the heaterbase, sensor cartridge, or expiratory heater wire adapter.
- Do not spray liquid into the vents or onto electrical connectors. Failure to comply may result in irreparable damage to the humidifier.
- Follow the responsible organization's guidelines for frequency of cleaning, rinsing, drying, handling and storage between uses.

### Routine maintenance

A full technical description, including routine maintenance and service data, is contained in the Product Technical Manual available from your supplier or Fisher & Paykel Healthcare.



**WARNING:** The Product Technical Manual must be followed for all servicing and maintenance of the humidifier. Failure to comply may impair performance of the humidifier or compromise safety (including potentially causing serious harm).

## Warnings, cautions and notes

---



### WARNINGS

- Refer to the instructions for use for breathing circuits, interfaces and accessories before operating the equipment. Failure to comply may impair performance of the humidifier or compromise safety (including potentially causing patient harm).
- This product is only designed and verified for use with accessories and spare parts approved by Fisher & Paykel Healthcare. Unauthorized accessories or spare parts which are used with the humidifier may impair performance of the humidifier, or compromise safety (including potentially causing serious patient harm), or result in increased electromagnetic emissions, or decreased electromagnetic immunity, resulting in improper operation.
- This product is designed for the delivery of air and/or oxygen. It is not suitable for the delivery of flammable anesthetic gas mixes or Heliox gas. Failure to comply may impair performance of the humidifier or compromise safety (including potentially causing patient harm).
- The humidifier should always be level and positioned lower than the patient. Failure to comply may impair performance of the humidifier or compromise safety (including potentially causing serious patient harm).
- Visually inspect components and accessories for damage before use and replace if damaged. Use of damaged components or accessories (including degraded sensors) may impair performance of the humidifier or compromise safety (including potentially causing serious harm).
- Appropriate patient monitoring (e.g. oxygen saturation) must be used at all times. Failure to monitor the patient (e.g. in the event of an interruption to gas flow) may result in serious harm or death.
- Do not touch the electrical connectors and the patient simultaneously. Failure to comply may result in serious harm.
- Operation of the humidifier outside of the recommended operating conditions (as described in these user instructions) may impair performance of the humidifier or compromise safety (including potentially causing patient harm).
- Monitor circuit condensate every six hours to prevent occlusion or build-up of fluid. Drain as required. Failure to comply may impair performance of the humidifier or compromise safety (including potentially causing serious patient harm).
- Remove any sources of ignition, such as: cigarettes, an open flame, or materials which ignite easily at high oxygen concentrations.
- Follow the instructions of the oxygen device provider; keep oxygen regulators, cylinder valves, tubing, connections, and all other oxygen equipment away from oil, grease, or greasy substances. Spontaneous and violent ignition may occur if these substances come into contact with oxygen under pressure.

## Warnings, cautions and notes

---

- California residents, please be advised of the following, pursuant to Proposition 65: This product contains chemicals known to the State of California to cause cancer, birth defects and other reproductive harm. For more information, please visit: <http://www.fphcare.com/prop65>
- The operation of high-frequency surgical apparatus, shortwave or microwave equipment in the vicinity of the humidifier may adversely affect its performance. If this occurs, remove the humidifier from the vicinity of such devices.
- Do not use this product in or near a magnetic resonance imaging (MRI) scanner, where the intensity of electromagnetic disturbances is high. Failure to comply may impair the performance of the humidifier or compromise safety (including potentially causing serious patient harm).
- Do not connect the humidifier directly to a medical gas pipeline system. The humidifier is intended for connection to a ventilator or gas mixer to control gas pressure and flow rate. Failure to control the gas delivery may result in a pressure injury to the patient.
- Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, observe all equipment to confirm that it is operating normally.
- The F&P 950 Respiratory Humidifier and accessories contain small parts which could cause injury or suffocation if inhaled or swallowed.
- Install the humidifier away from heat sources, such as direct sunlight, radiant heaters, fireplaces, ovens and kettles and cooling sources, such as dehumidifiers, fans, air conditioners and ventilators. Failure to comply may impair the performance of the humidifier or result in serious harm.
- To avoid strangulation or tripping, ensure the breathing tubes and power cord are positioned in a tidy manner away from the floor and patient, so they will not get entangled or wrapped around the limbs or neck.
- Consideration should be given to the possible hazards that could arise from children, pests and pets.



### CAUTIONS

- Ensure that Invasive mode is set for patients who have bypassed airways. Prolonged exposure to reduced humidity will result in patient harm including decreased mucociliary clearance, atelectasis, or pneumonia.
- Do not touch the hot surface of the heater plate, chamber base or probes. Failure to comply may result in a skin burn.
- The F&P 950 Respiratory Humidifier does not contain material known to cause allergic reactions. If an allergic reaction occurs during use, contact the responsible organization immediately.

### NOTES

- Use USP sterile water for irrigation, or equivalent. Adding other substances may have adverse effects.
- The F&P 950 Respiratory Humidifier contains an embedded software system licensed to Fisher & Paykel Healthcare by Microsoft. The license contains certain restrictions that are relevant to the use of the F&P 950 Respiratory Humidifier. Visit [www.fphcare.com/microsoftlicensing](http://www.fphcare.com/microsoftlicensing) for more information about such restrictions.
- The F&P 950 Respiratory Humidifier has an IP21 rating, which protects against solid foreign objects 12.5 mm in diameter and uniform flow of water drops over the enclosure area with a flow rate of 1 mm/min.
- This equipment's emissions characteristics make it suitable for use in industrial areas and hospitals (CISPR 11 class A) and residential environments (CISPR 11 class B).
- If a serious incident has occurred while using this device, please inform your local Fisher & Paykel Healthcare representative and, for European Union member countries, the Competent Authority in your country.

## Symbol definitions

 Follow instructions for use - safety	 Consult instructions for use. <a href="http://www.fphcare.com/950IFU">www.fphcare.com/950IFU</a>	 Manufacturer	 Date of manufacture	 Catalogue reference number	 Batch code	 Serial number
 Type BF Applied Part	 Class II equipment	 Alternating current	 Standby (On/Off)	 IP Classification	 Temperature limitations	 Humidity limitations
 USB 2.0	 WEEE (Waste Electrical and Electronic Equipment)*	 European representative*	 CE Marking 93/42/EEC*	 Regulatory Compliance Mark*	 Raise finger guard	 Date of expiration
 Fragile, handle with care	 Keep dry	 Recyclable	 Caution	 Warning	 Alarm	 Menu
 Warning: hot surface	 Alarm audible pause	 Alarm audible paused	 Invasive mode	 Mask mode	 Optiflow mode	 Neonatal mode
 Neonatal Invasive mode	 Neonatal CPAP   NIV mode	 Neonatal Optiflow mode	 Sensor cartridge authentication failure	 Sensor Cartridge service life warning	 Accept	 Cancel
 Back arrow	 Locked	 Unlock	 Medical Device*	 Prescription only*	 Importer	 Distributor
 Authorized representative for Switzerland*	 UK responsible person*	 INMETRO Mark*		 UL Mark*		

\*symbol displayed on select models

## Technical specifications

### Product specifications

	Heaterbase Specifications		
<b>Dimensions (heaterbase only)</b>	240 mm (D) x 154 mm (W) x 253 mm (H)		
<b>Weight (heaterbase and power cord only)</b>	3.45 kg		
<b>Supply frequency</b>	50/60 Hz		
<b>Supply voltage</b>	<small>REF</small> <b>950AXX</b> <sup>1</sup> 230 V <small>REF</small> <b>950JXX</b> <sup>1</sup> 115 V <small>REF</small> <b>950GXX</b> <sup>1</sup> 100 V		
<b>Power rating</b>	350 VA		
<b>Maximum length of power cord</b>	3.3 m		
<b>Sound pressure level</b>	Alarms exceed 45 dbA @ 1 m		
<b>Auditory alarm pause</b>	120 seconds		
<b>Maximum temperature of delivered gas</b>	43 °C		
<b>Time to reach set temperature (gas flow is required)</b>	< 30 minutes		
<b>Maximum surface temperature of the breathing circuit (applied part section)</b>	44 °C		
<b>Component service life</b>	Heaterbase: 7 years		
	Adult	Pediatric	Neonatal
<b>Humidity performance (Except in the event of a humidifier alarm or power failure or electromagnetic disturbance)</b>	<b>Invasive mode:</b> > 33 mg/L <b>Mask mode:</b> > 12 mg/L <b>Optiflow mode:</b> > 12 mg/L	<b>Invasive mode:</b> > 33 mg/L <b>Mask mode:</b> > 12 mg/L <b>Optiflow mode:</b> > 12 mg/L	<b>Neonatal Mode:</b> > 33 mg/L <b>Invasive mode:</b> > 33 mg/L <b>CPAP   NIV mode:</b> > 12 mg/L <b>Optiflow mode:</b> > 12 mg/L
<b>Operating flow range (L/min, STPD)</b>	<b>Invasive mode:</b> 5-60 L/min <b>Mask mode:</b> 5-120 L/min <b>Optiflow mode:</b> 5-70 L/min	<b>Invasive mode:</b> 1-60 L/min <b>Mask mode:</b> 1-60 L/min <b>Optiflow mode:</b> 1-60 L/min	<b>Neonatal Mode:</b> 0.5-40 L/min <b>Invasive mode:</b> 0.5-40 L/min <b>CPAP   NIV mode:</b> 0.5-40 L/min <b>Optiflow mode:</b> 0.5-36 L/min

<sup>1</sup>XX represents the country code

## Technical specifications

### Operating conditions

SPECIFICATION	ADULT	PEDIATRIC & NEONATAL
Room temperature	18–26 °C	20–26 °C
Incoming gas temperature	Minimum = Room temperature Maximum = 10 °C above room temperature (at 30% relative humidity)	Minimum = Room temperature Maximum = 10 °C above room temperature (at 30% relative humidity)
Operator position	<1 m from heaterbase	<1 m from heaterbase
Atmospheric pressure:	Minimum of 70 kPa (equivalent to a maximum altitude of 3000 m) Maximum 106 kPa	Minimum of 70 kPa (equivalent to a maximum altitude of 3000 m) Maximum 106 kPa

### Storage conditions

SPECIFICATION	VALUE
Temperature	-20–60 °C
Humidity	10–95% relative humidity non-condensing

**NOTE:** If the humidification system has been stored outside the specified operating ambient temperature range, the system must be left for 24 hours within the specified operating temperature range before use.

### Disposal

At the end of its life, users should dispose of the humidifier according to the responsible organization's guidelines, local authority guidelines, and national electrical and electronic equipment regulations.







Fisher & Paykel Healthcare Ltd, 15 Maurice Paykel Place, East Tamaki, Auckland 2013, PO Box 14 348 Panmure, Auckland 1741, New Zealand Tel: +64 9 574 0100 Fax: +64 9 574 0158 Email: info@fphcare.co.nz Web: www.fphcare.com



**Australia (AU)** (Sponsor) Fisher & Paykel Healthcare Pty Ltd, 19-31 King Street, Nunawading, Melbourne, Victoria 3131. Tel: +61 3 9871 4900 **Brazil (BR)** Fisher & Paykel do Brasil, Rua Sampaio Viana, 277 cj 21, Paraíso, 04004-000, São Paulo – SP, Brazil Tel: +55 11 2548 7002 **China (CN)** 代理人/售后服务机构: 费雪派克医疗保健 (广州) 有限公司. 广州高新技术产业开发区科学城科丰路31号G12栋301号 电话: +86 20 32053486 **France (FR)** Fisher & Paykel Healthcare SAS, 10 Av. du Québec, Bât F5, BP 512, Villebon-sur-Yvette, 91946 Courtaboeuf Cedex, France Tel: +33 1 6446 5201 Email: c.s@fphcare.fr **Germany (DE)** Fisher & Paykel Healthcare GmbH, Wiesenstrasse 49, 73614 Schorndorf, Germany Tel: +49 7181 98599 0 **Hong Kong (HK)** Tel: +852 2116 0032 **India (IN)** Tel: +91 80 2309 6400 **Japan (JP)** Tel: +81 3 5117 7110 Fax: +81 3 5117 7115 **Korea (KR)** Tel: +82 2 6205 6900 **Mexico (MX)** Tel: +52 55 9130 1626 **Poland (PL)** Fisher & Paykel Healthcare Poland Sp. z o.o., Pl. Andersa 7, 61-894 Poznań, Poland Tel: +48 664 846 464 **Russia (RU)** Tel: +7 495 782 21 50 **Switzerland (CH)** Fisher & Paykel Healthcare GmbH, Säntisstrasse 2, 9501 Wil / SG, Switzerland Tel: 0800 83 47 63 **Taiwan (TW)** Tel: +886 2 8751 1739 **Turkey (TK)** İthalatçı Firma: Fisher Paykel Sağlık Ürünleri Ticaret Limited Şirketi, İletişim Bilgileri: Ostim Mahallesi 1249. Cadde No:6, Yenimahalle, Ankara, Türkiye 06374, Tel: +90 312 354 34 12 **UK** Fisher & Paykel Healthcare Ltd, Unit 16, Cordwallis Park, Clivemont Road, Maidenhead, Berkshire SL6 7BU, UK Tel: 0800 132 189 **USA (US)/Canada (CA)** Tel: 1800 446 3908 or +1 949 453 4000

**Austria (AT)** Tel: 0800 29 31 23 **Benelux (BE NL LU)** Tel: +31 40 216 3555 **Denmark (DK)** Tel: +45 70 26 37 70 **Finland (FI)** Tel: +358 9 251 66 123 **Ireland (IE)** Tel: 1800 409 011 **Italy (IT)** Tel: +39 06 7839 2939 **Norway (NO)** Tel: +47 21 60 13 53 **Spain (ES)** Tel: +34 902 013 346 **Sweden (SE)** Tel: +46 8 564 76 680

F&P, F&P 950 and Optiflow are trademarks of Fisher & Paykel Healthcare Limited.  
For patent information, see www.fphcare.com/ip

0123