



AQUA MEASURE INSTRUMENT CO. & MOISTURE REGISTER PRODUCTS DIVISION

1712 Earhart Court • La Verne, CA 91750
TEL: (909) 392-5833/36 • FAX: (909) 392-5838 • e-mail: aquamoist@aol.com
www.moistureregisterproducts.com

CERTIFIED TEST REPORT

DATE: 7 / 07 / 06

MODEL: AM-1

SERIAL NO.: 1351

CURVE: 3

POSITION: 12:00

ORDER NO. MP23606

CUSTOMER: ABITIBI CONSOLODATED

ADDRESS: 17 MI PO BOX 128 SNOW FLAKE, AZ

PURCHASE ORDER NO. _____

STANDARDS USED: PHYSICAL DIELECTRIC STANDARDS - (PRIMARY)

The AM-1 Instrument is calibrated to Physical Dielectric Standards which are materials with fixed and stable dielectric properties. These Standards represent products of various moisture contents to the AM-1 Instrument.

CALIBRATION DATA

PHYSICAL STANDARD	STD. TOL. CURVE 3	AM-1 READING CURVE 3	STD. TOL. CURVE	AM-1 READING CURVE
BAKELITE	11.0 ± 0.1	11.0	±	
PHENOLIC	10.5 ± 0.1	10.5	±	
NYLON	6.6 (+0.1, -0.2)	6.8	±	
H.D. POLY	0.9 ± 0.2	0.7	±	
INTERNAL STD. (1)	8.2 ± 0.1	8.3	±	
FULL SCALE	14.9 ±	14.9	±	
ROLLERS	CHECK OK			

AQUA MEASURE INSTRUMENT COMPANY CERTIFIES THAT THIS INSTRUMENT HAS BEEN INSPECTED AND TESTED TO COMPLY IN EVERY WAY WITH THE RELEVANT SPECIFICATIONS AND OR DRAWINGS, AND IN ACCORDANCE WITH REQUIREMENTS OF THE REFERENCED PURCHASE ORDER.

Manufactured and Assembled in the United States of America.

SIGNED:

Todd Bluhm

Certified Test Report

AMI3.DOT

Revision: C

March 17, 2003

P/N: 552662-2

INSTRUMENT TECHNICIAN