


Compass 315M

CW Green Lasers



Features

- Highest performance and output power
- Best laser beam parameters
- Proven reliability
- Highest unit-to-unit consistency
- Unsurpassed lifetime



Compass™ 315M

CW Green Lasers

Specifications	Wavelength (nm)	532			
	Output Power¹ (mW)	20	50	100	150
	Spatial Mode	TEM _{OO}			
	Circularity of Beam	>95%, <1.1:1	>95%, <1.1:1	>95%, <1.1:1	>90%, <1.15:1
	Beam Diameter @1/e² (mm)	0.32	0.32	0.32	0.34
	Beam Divergence (mrad)	⟨2.2			
	Pointing Stability (µrad/°C)	<6			
	Noise (10 Hz to 1 GHz)(rms)	<0.25%			
	Long-Term Power Stability (over 8 hours)	<±2%			
	Warm-∪p Time (min.)	< 10			
	Polarization Ratio	100:1, vertical			
	Static Alignment Tolerances ²				
	Beam Position (mm)	±0.1			
	Beam Angle (mrad)	±0.5			
Environmental Specifications	Operating Voltage³ (VDC)	+12 to +28			
	Power Consumption (W)	<150 (typically 40)			
	Maximum Laser Head Baseplate Temperature				
	for Conductive Cooling	45°C (113°F) ⁴			
	Storage Temperature	-40°C to 80°C (-40°F to 176°F)			
	Dimensions (L x W x H)				
	Laser Head	100 x 40 x 31.5 mm (4.0 x 1.6 x 1.3 in.)			
	Analog Controller	100 x 129 x 41 mm (4.0 x 5.1 x 1.6 in.)			
	Heatsink (optional) Power Supply (optional)	155 x 58 x 40 mm (6.1 x 2.3 x 1.6 in.)			
	Cable—Laser Head to Controller	255 × 126 × 36 mm (10 × 4.6 × 1.4 in.) 2 m (6.6 ft.)			
	Weights	2 111 (0.0 11.)			
	Laser Head	0.33 kg (0.73 l	he)		
	Analog Controller	0.38 kg (0.83 lbs.)			
	Heatsink (optional)	0.50 kg (1.1 lbs.)			
	Power Supply (optional)	1.66 kg (3.66 lbs.)			
	Cable—Laser Head to Controller	0.52 kg (1.15 lbs.)			
Measurement Tools	Detector	FieldMax-TO			
	Sensor	OP-2VIS			
	Output power is variable via analog controller or optional digital controller. Specifications are valid for nominal power. Recommended power range is 60-100% of nominal power. Static alignment tolerances are relative to fitting holes in base plate. The power supply must meet the following requirements. DC Voltage +12V to +28V Power 150W Ripple <5% peak-to-peak For Compass 315M-150 only.				

Coherent follows a policy of continuous product improvement. Specifications are subject to change without notice.

Coherent's scientific and industrial lasers are certified to comply with the Federal Regulations (21 CFR Subchapter J) as administered by the Center for Devices and Radiological Health on all systems ordered for shipment after August 2, 1976.

Coherent offers a limited warranty for all Compass lasers. For full details of this warranty coverage, please refer to the Service section at www.Coherent.com or contact your local Sales or Service Representative.



Coherent, Inc.

5100 Patrick Henry Drive Santa Clara, CA 95054 phone (800) 527-3786 (408) 764-4983

(408) 764-4983 fax (800) 362-1170 (408) 988-6838

(408) 988-6838 e-mail tech.sales@Coherent.com Benelux +31 (30) 280 6060 China +86 (10) 6280 0209 France +33 (0)1 6985 5145 Germany +49 (6071) 968333 Italy +39 (02) 34 530 214

Haly +39 (02) 34 530 214 Japan +81 (3) 5635 8700 Korea +82 (2) 460 7900 UK +44 (1353) 658 833

