

# OmniCure® LX500 Series

## UV Curing Solutions

Applicable for the following UV LED Head Controller models:

Model	Part Numbers
LX500-2 (supports two LED Heads)	010-00369R
LX500-4 (supports four LED Heads)	010-00375R

## Installation/Reference Guide

Excelitas Canada Inc.  
2260 Argentia Road  
Mississauga (ON)  
L5N 6H7 Canada  
+1 905.821.2600  
[www.excelitas.com](http://www.excelitas.com)

035-00628R Rev 0

# OmniCure® LX500 Series

UV Curing Solutions

---

## Installation/Reference Guide

035-00628R

Excelitas Canada Inc. 2016

All rights reserved

No part of this publication may be reproduced, transmitted, transcribed, stored in a retrieval system or translated into any language in and form by any means without the prior written consent of Excelitas Canada Inc. every effort has been made to ensure information in this manual is accurate, however, information in this manual is subject to change without notice and does not represent a commitment on the part of the authors.

## Trademarks:

OmniCure® is a trademark of Excelitas Canada Inc. All rights reserved. All other product names are trademarks of their respective owners. All product or software photos shown are for reference only and are subjected to change without notice.

Made in Canada. Doc. No. 035-00628R

---

<http://www.excelitas.com/omnicure>

Excelitas Canada Inc. 2016

All rights reserved

ii

035-00628R

## OmniCure® LX500 Series

UV Curing Solutions

---

### Revision History

Revision #	Comments
Rev 0	Initial Release

## Table of Contents

<b>1</b>	<b>Introduction .....</b>	<b>1</b>
<b>2</b>	<b>Safety Precautions/ User Warnings .....</b>	<b>2</b>
2.1	Glossary of Symbols .....	2
2.2	Safety Precautions .....	2
<b>3</b>	<b>Optical Safety Data.....</b>	<b>4</b>
<b>4</b>	<b>Getting Started .....</b>	<b>5</b>
4.1	Package Contents .....	5
4.2	UV LED Controller and Head Images .....	6
<b>5</b>	<b>Installing the Hardware .....</b>	<b>7</b>
5.1	General Installation .....	7
5.2	Clamp Installation and Removal .....	7
5.3	Changing Lenses .....	8
5.4	Storing LED Head and Lens Assemblies.....	8
5.5	Installation Component Images.....	8
<b>6</b>	<b>Operation.....</b>	<b>10</b>
6.1	Powering Up and Powering Down .....	10
6.2	Understanding the Controller Display.....	11
6.3	Closed Loop Feedback .....	14
6.4	Defining Exposure Parameters .....	14
6.5	Running an Exposure from the Controller .....	17
6.6	Using the PC Interface with USB Connection .....	17
6.7	Interfacing with a PLC .....	17
6.8	Using the Foot Pedal .....	19
<b>7</b>	<b>Troubleshooting .....</b>	<b>21</b>
7.1	Symptoms and Possible Causes.....	21
7.2	Error Codes .....	22
<b>8</b>	<b>Care and Maintenance.....</b>	<b>24</b>
8.1	Cautions and Handling .....	24
8.2	Cleaning Materials.....	24
8.3	Cleaning the LED Head and Lens Assembly .....	25
8.4	Cleaning the Controller .....	25
8.5	Calibrating a UV LED Head.....	25
8.6	Viewing Log Files .....	26
8.7	SD Cards .....	26
<b>9</b>	<b>Technical Specifications .....</b>	<b>30</b>
9.1	Environmental Conditions .....	30
9.2	Controller Specifications .....	30
9.3	UV LED Head Specifications.....	31
<b>10</b>	<b>Regulatory Compliance .....</b>	<b>32</b>
10.1	Product Safety and Electromagnetic Compatibility:.....	32
10.2	China RoHS.....	32
10.3	WEEE Directive (2012/19/EU) .....	33

<b>11</b>	<b>Warranty.....</b>	<b>34</b>
11.1	Checking Your Warranty Status.....	34
11.2	Replacement UV LED Module Warranty .....	34
11.3	Returning your LX500 System to Excelitas Technologies for Service .....	34
<b>12</b>	<b>Contact Information .....</b>	<b>36</b>

## Figures

Figure 1	UV LED Head Controller (front) .....	6
Figure 2	UV LED Head Controller (back) .....	6
Figure 3	UV LED Head .....	6
Figure 4	Lens/LED Head Connection.....	8
Figure 5	Recommended Clamp Position .....	9
Figure 6	Typical Installation Applications .....	9
Figure 7	Extension Rod Application.....	9
Figure 8	PLC Connector (looking at Controller).....	17
Figure 9	PLC digital output equivalent circuit .....	19

## Tables

Table 1	PLC Connector Pin Out .....	19
Table 2	Error Codes.....	23
Table 3	Environmental Conditions .....	30
Table 4:	Controller Specifications .....	31

### 1 Introduction

The OmniCure® LX500 series represents a new standard in UV LED Curing. It gives you the power, control, and reliability never before available in such a form factor, providing configuration flexibility to suit any application for end users and integrators alike.

The UV LED Head Controller allows optical output to be controlled manually via touch pad control, or by PLC or PC (Virtual COM Port) USB interface. The UV LEDs are enclosed within the UV LED Head which provides excellent spectral and power output. The UV LED Head has a typical lifetime in excess of 20,000 hours of accumulated 'on' time at either 365nm or 385nm wavelengths.

The output power emission of the available UV LED Head is adjustable in 1% increments from 5% to 100% to provide light for your application. In order to set your OmniCure LX500 series at specific irradiance levels (W/cm<sup>2</sup>), we suggest running trials to optimize the results for the UV cured glues and adhesives used.

In the Excelitas Technologies family of UV LED light systems, all systems offer the same high level of innovation, quality and reliability that our customers have come to expect. OmniCure® has combined next generation optical engineering, state-of-the art electronics and fibre-optics to produce sophisticated technologies that employ light. Today, OmniCure® is a leading developer of light-based systems for sectors ranging from manufacturing to bio-medicine, and is unmatched in commitment to quality and service.

This manual covers the following models:

LX500-2	010-00369R
LX500-4	010-00375R

Excelitas Technologies recommends reading this guide to discover all features of the OmniCure® LX500 series systems, and how to use them.

**Note:** Refer to OmniCure® LED Head Assembly Specification Guide 035-00638R for LED Head specifications.

## 2 Safety Precautions/ User Warnings

### 2.1 Glossary of Symbols



Caution risk of danger – consult accompanying documents.



Risk Group

WARNING: UV emitted from this product. Avoid eye and skin exposure to unshielded product.

WARNING: Possibly hazardous optical radiation emitted from this product. Do not stare into operating lamp.



Input/ Output Signals



Input Signal



Caution, Hot Surface



CAUTION, Risk of Electrical Shock

### 2.2 Safety Precautions

This series of cautions and warnings relate to the installation, operation and maintenance of the OmniCure® LX500System. They are also presented throughout this Installation/Reference Guide as applicable.



WARNING

Do not stare directly at the light emitted from the LED array. This may be harmful, resulting in eye injury.

Always use UV protective eyewear as indicated below. Additionally, protect any exposed skin with appropriate clothing or shielding as required.



Warning: UV protective eyewear must meet the following recommended optical specifications:

- Spectral range; 350-440nm
- Optical Density  $\geq 6$



Warning: Should the OmniCure® LX500 System be used in a manner not specified by Excelitas Technologies, the protection provided by the equipment may be impaired.



The UV LED Head operation can be affected if handled incorrectly. Never touch the protective lens cover assembly of the UV LED Head. The presence of skin oils may result in a decrease in system performance.



Always ensure the power supply is turned off prior to disconnecting or re-connecting the UV LED Head.

## OmniCure® LX500 Series

UV Curing Solutions

---



Warning:

Monitoring the system during automated operation:

The level of UV energy supplied by the OmniCure® LX500 System is sufficient to ignite flammable substances. Therefore, when the unit is operated unattended in an automated environment, an alarm function must be provided by the user to indicate a malfunction in the associated equipment used. During installation the appropriate measures must be included to prevent any such occurrence.



Warning:

Always use this product with the external power supply and power cord set provided. Ensure the external power supply cord is connected to a grounded 3-pin outlet only. Any substitution of these components will invalidate regulatory certification of this product and may impair operating safety.



### 3 Optical Safety Data

IEC 62471: Photobiological Safety of Lamps and Lamp Systems

Resulting Classification and Labelling



## 4 Getting Started

### 4.1 Package Contents

Your package contains the following:

- UV LED Head Controller (2-head 010-00369R or 4-head 010-00375R)
- Door lock jumper (018-00648R)
- DC power supply (290-00041R)
- UV protective eyewear (854-00001R)

Carefully unpack the unit and store the packing material for future use.

#### Optional Components

In addition to the Controller, you will require one to four LED Heads and lenses, depending on your configuration. Other optional components are available that may be sourced separately or may be purchased from Excelitas Technologies. Check with Excelitas Technologies to determine the appropriate components for your requirements. The following table provides the part numbers of component parts:

Item	Part Number
LED Heads	
365nm x 55mm LED UV Head, Inc Clamp Sub Assembly	019-00286R
365nm x 125mm LED UV Head, Clamp Assembly Not included	019-00287R
385nm x 55mm LED UV Head, Inc Clamp Sub Assembly	019-00288R
385nm x 125mm LED UV Head, Clamp Assembly Not included	019-00289R
Lenses	
3mm focusing lens for 365nm/385nm	810-00053R
6mm focusing lens for 365nm/385nm	810-00054R
8mm focusing lens for 365nm/385nm	810-00060R
10mm focusing lens for 365nm/385nm	810-00061R
12mm focusing lens for 365nm/385nm	810-00066R
5mm Cylindrical Lens for 365nm/385nm	810-00078R
90 deg Adapter, 6mm spot, 365nm/385nm	810-00083R
90 deg Adapter, 8mm spot, 365nm/385nm	810-00084R
90 deg Adapter, 10mm spot, 365nm/385nm	810-00085R
Extension Cables	
Extension cables: 1m	018-00642R
Extension cables: 3m	018-00643R
Extension cables: 5m	018-00644R
Extension cables: 10m	018-00645R
Accessories	
Mounting clamp?	019-00288R
Foot pedal switch assembly	014-00070R
LS100 Light Sensor	019-00237R

A micro SD card is useful (but not supplied by Excelitas) to save log files for monitoring or to save settings files for future re-use.

## OmniCure® LX500 Series

UV Curing Solutions

### 4.2 UV LED Controller and Head Images

Note: LX500-4 shown.




Figure 1 UV LED Head Controller (front)




Figure 2 UV LED Head Controller (back)



Figure 3 UV LED Head

## 5 Installing the Hardware

### 5.1 General Installation

1. Determine the desired UV LED Head location and position based on the results of a series of testing and trials.
2. Determine the working location of the Controller, ensuring access to the AC power source does not cause any possible obstruction with the power cable connection from the source to the controller. The main power switch is located on the front panel of the unit. The unit should be placed so that the switch is convenient for the operator to access.
3. If you are not integrating with an existing controller, ensure the door lock jumper is connected.
4. Connect the UV LED Head cable to the rear of the Controller. Only use Excelitas extension cables to extend the reach of the UV LED Head.
5. If using a foot pedal switch, connect it to the corresponding input on Controller (e.g. Head 1, FP 1). Refer to Figure 2.
6. Install the supplied installation clamp to the existing fixture using either the 3 metric threaded screw holes or the 3 imperial threaded screw holes. The two end faces of the installation bracket are identified for metric and imperial thread sizes. Though the clamp can be located in any position on the head, for optimum heat dissipation resulting in optimum performance it is recommended that the clamp position on the head is as per Figure 5.
7. Install the UV LED Head into the installation clamp and secure its location/position by tightening the hex screw. Refer to Figure 6 for typical installation applications.
8. Attach the appropriate lens and remove the protective cap. To remove the cap, pull the cap-hat — do not rotate the LED head or cap.  
When not in use, reinstall the cap by holding the cap-hat and pushing it back over the head. This will prevent any contamination to the lens when the head is not used for more than one day. Refer to Figure 4.
9. If you are not integrating with an existing controller, connect the door lock jumper between pins 24 and 25 (refer to Table 1). This jumper can be connected to your UV protective enclosure to prevent accidental UV exposure during operation.  
Closed contact – enable UV LED Head(s) output  
Open contact – disable UV LED Head(s) output.
10. Connect the AC power cord in to the power supply and to a properly grounded AC outlet. For safe operation, only use the power cord supplied or one with an equivalent rating, shield, and approvals.
11. Connect the power supplies DC barrel connector to the back of the Controller.

**Note:** The UV LED Head connectors are locking. To disconnect them from the Controller, squeeze the tab on the plastic shell.

### 5.2 Clamp Installation and Removal

1. Ensure the clamp locking screw (M4 x 10, Allen key size 3mm) is loose before slipping the clamp into the LED lens and head assembly.
2. Slide the clamp over the lens housing and onto the head and tighten the locking screw. For optimal performance and ease of changing lenses, position the clamp so the bottom face is located between the lens housing crease and the reference line (first circular groove) on the head body. Refer to Figure 4 and Figure 5.

3. To remove the clamp loosen the locking screw and slide the clamp off the lens and head assembly.

### 5.3 Changing Lenses

*Do not use any mechanical tool when removing or replacing the lens/housing.  
Hand tighten only, do not over tighten as this will make the assembly difficult to remove.  
The lens and housing are one subassembly. Do not attempt to remove the lens from the housing.  
Ensure the system is off and the head has cooled to room temperature before changing the lens.*

Holding the head firmly unscrew the lens housing and replace with desired replacement lens. Refer to [Figure 4](#).

It is recommended to clean the replacement lens thoroughly prior to replacement. Refer to [Section 8.3 Cleaning the LED head and Lens assembly](#). When not in use, place protective cap over the lens.

### 5.4 Storing LED Head and Lens Assemblies

When not in use, store the head and lens assembly individually wrapped in a clean, lint free bag in a safe, dry place.

Never store multiple unwrapped head and lens assembly in the same bag, as they may become damaged if they come in contact with each other.

### 5.5 Installation Component Images




Figure 4 Lens/LED Head Connection




Figure 5 Recommended Clamp Position




Figure 6 Typical Installation Applications



Figure 7 Extension Rod Application

## 6 Operation

### 6.1 Powering Up and Powering Down



**CAUTION:**

Ensure that the Controller is turned "OFF" before connecting or disconnecting UV LED Heads.

**WARNING:**

Before running an exposure, make sure all operators in the area are wearing appropriate eye protection or the UV LED Heads are behind a UV blocking enclosure.


1. Ensure that the UV LED Head is securely mounted in the desired orientation, the UV LED Head is connected to the back of the Controller, and the PLC remote door lock jumper is properly installed. Refer to Figure 2 and Section 5, "Installing the Hardware".

2. Verify that the Controller power supply is connected to a properly grounded AC outlet.



Warning – Do not use inadequate detachable power cords. Please refer power technical specification for more detailed information

3. Turn the Controller power switch to "ON" (refer to Figure 1).
4. The LCD screen will turn on and the Excelitas Technologies logo and the software version number will appear. A warning message may appear indicating that the Head is not in the same unit/head port position it was used in previously. At this message press **Select** to clear the warning message or power off and correct the connections.
5. When the title screen for the LX500 disappears and the Main Control Panel is displayed, the system is ready for use. The Main Control Panel shows the current state of any connected UV LED Heads.



Caution – the UV LED Head can become very hot; avoid handling until the UV LED Head has cooled (approx. 5 min) or use protective gloves.

Before connecting or disconnecting any UV LED Heads, ensure that the Controller is turned "OFF".

To power down the unit, turn the Controller power switch to "OFF".

**Note :** The main power switch is located on the front panel. The unit should be placed so the switch is convenient for the operator to access.

### 6.2 Understanding the Controller Display

---

#### Status Bar



The top left corner of the display shows the menu icon, which allows you to switch between screens. When nothing else on the screen is selected, the menu icon is. The status bar may also display the following icons:



– indicates that the PLC door pin is open and LED exposures have been disabled.




– indicates that the front panel interface has been locked out. To lock the front panel, use the PLC inputs or PC command. (Lock status is not stored between power cycles)



(up or down arrow) – indicates that the unit is in count up or countdown mode.

#### Menu



Press **Select** at any time to display the menu.


Use the arrows to change selection, then press **Select**.



## OmniCure® LX500 Series

UV Curing Solutions

### Main Control Panel



Displays upon start-up; shows current state of each connected UV LED Head (displays two or four channels depending on Controller model).

The timer for each channel is how long (in seconds) the UV LED Head will remain on when an exposure is triggered. If the timer is displaying 0.0 seconds then the unit is in "count up" mode and the UV LED Head will remain on for as long as the enable signal is present. If in count up mode, the timer value cannot be changed.

The intensity is represented by both a numerical value and a bar graph. If calibrated for irradiance, the number will be a decimal without a 'W', if a decimal number with a W is shown, the head is calibrated for power mode. The color of the intensity bar will be same as the color code on the UV LED Head. The intensity will be displayed as a percentage for un-calibrated heads or in W/cm<sup>2</sup> or W for calibrated heads (refer to [Section 8.5, "Calibrating a UV LED Head"](#)).

Closed loop feedback will be engaged for intensities below 100%. When closed loop feedback (CLF) is on, an icon will appear beside the intensity bar for that channel. The CLF icon appears when CLF is actively controlling the output of the head. If the CLF icon disappears before the completion of an exposure, then CLF has reached the upper limit it can adjust for and is no longer functional. If CLF usage is desired, it is recommended that you use an intensity of 80% or less.

The bottom of the screen shows the last event. If you have an SD card installed this event will be saved to a log file. (refer to [Section 8.6, "Viewing Log Files"](#)).

When the system experiences an alarm, the alarm icon beside the Head number will turn red and an alarm message will pop up.

### StepCure

Mar 28, 2016 10:05					
	#	Time	Level	Dwell	Src
1	1	1.0	75	0.1	FP1
2	2	0.5	71	0.1	FP1
3	3	0.5	69	0.1	FP1
4	R	Repeat 2 times			FP1
5					
6					
7					
8					

Use to control the UV LED Head operation for sequence programming and S2000 compatibility. For more information, see [Section 6.4, "Defining Exposure Parameters"](#).

### Settings

	Mar 28, 2016 10:05
Time and Date	Mar 28, 2016 10:06
PLC Delay Timing	0ms
PLC Logic	Normal
Language	English

Displays current system statistics. If necessary, use the arrow keys to select and change any field.

- *Time and Date*
- PLC Delay Timing and Head association – add a positive or negative delay to the Programmable Sync Out pin. By default, the PLC delay timing is off, pressing the right arrow button will flash the head the PLC delay is to be associated with. Press up or down to change the head association. The PLC delay timing is adjustable in 10ms increments from 0ms to +/-30,000ms.
- *PLC Logic* – invert the PLC input and output signal logic. Only the door lock logic is not inverted. **Note:** When changing, the interface will lock due to the signal flip. Ensure you can control the PLC lock signal before toggling this setting. Can be corrected via PC interface.
- *Language* – switch the language displayed on the LCD screen. Currently only English is supported.


### LED Head Information

	1	2	3	4
WL	385	0	365	385
Hours	13	0	0	5
Cal Hr	0	0	0	0
Temp C	24	25	24	24
Max C	75	0	75	75
Detect	OK	OK	OK	OK
Error	0	0	0	0
Mode	Cnt Dn	Cnt Dn	Cnt Dn	Cnt Dn

Displays current programming for each connected UV LED Head:


- *WL* – wavelength in nm
- *Hours* – accumulated operational hours
- *Cal Hr* – Number of calibrated hours remaining.
- *Temp* – present temperature in °C
- *Max* – maximum allowable temperature (in °C) before an alarm is triggered
- *Detect* – detects the presence of a UV LED Head in that channel:
  - *NC* – Not connected at startup
  - *Lost* – Head disconnected during use
  - *OK* – Head connected and communicating
- *Error* – Error code ("0" indicates no error). Refer to [Section 7.2, "Error Codes"](#).
- *Mode* – indicates if channel is in count up or countdown mode. Use the arrows to select and edit this field. All values will change simultaneously as the LX500 only support one count mode for all Heads.

#### Calibration




Used to initiate and display the calibration process. Horizontal access is the intensity percentage setting, vertical access automatically scales and is the corresponding irradiance in W/cm<sup>2</sup> or power in W. See [Section 8.5, "Calibrating a UV LED Head"](#).

#### SD Card Maintenance



Displays contents of inserted SD card. The LX500 uses three files: two log files (refer to [Section 8.6, "Viewing Log Files"](#)) and a settings file (refer to [Section 6.4, "Defining Exposure Parameters"](#)). Save allows the current settings to be saved to the default filename. When you select files you will see a menu depending on the type of file that allow you to load settings, update the firmware, or delete files. To open a directory, highlight it and press **Select**. If in a sub directory, press **Select** on .... to go up a level. Only five subdirectories deep is supported by the LX500. The Usage bar shows the capacity of the SD card; the middle number reflects the current free space on the card.

### 6.3 Closed Loop Feedback

 This symbol means that the LED is currently being closed loop feedback controlled. Every time an intensity adjustment is made, whether by front panel or PC control, a new set-point must be acquired. This occurs 500ms after the LED is turned on, and is used during every subsequent exposure. If your exposure timer is less than 500ms, this set-point will never be obtained. Once the intensity adjustment limit is reached, CLF will be disabled due to an inability to increase the output intensity. When this symbol is displayed, CLF is active on the associated channel.

### 6.4 Defining Exposure Parameters

#### Programming an exposure

Use the Main Control Panel to change the timer or intensity for each UV LED Head:

- Use the arrow keys to highlight the field you wish to change. The text in the highlighted field will turn red.
- Press **Select** to select the field you wish to change. Selected fields will begin to flash.
- Use the up and down arrow keys to change the value of the field. Holding down an arrow key will increase how fast the value changes.

#### Programming multiple steps or multiple heads

Use the StepCure screen to configure input signals to trigger a step profile of one or more heads. StepCure allows for:

## OmniCure® LX500 Series

### UV Curing Solutions

- Running a sequence of exposures on one head, at varying times and intensities.
- Running a sequence of exposures on multiple heads, at varying times and intensities.
- Running parallel exposures on multiple heads, each can be at a different time and intensity.
- Programming multiple programs for the same head from different trigger sources.
- Setting which head will be controlled by the S2000 command set.

**Note:** If no head is programmed to respond to the S2000 command set, sending S2000 commands will return errors.

**Note :** For customer interfaces designed around the S2000 lamp based system migrating to the LX5 system, the S2000 command set is compatible with the LX5 via the PC USB interface. Please contact Excelitas for more information regarding the S2000 Command Set.

**Note:** A StepCure program can only be run by using the trigger source specified. If another trigger source (not assigned in the StepCure screen to a program) is triggered, the head will run the profile specified on the main control screen.

**Note:** StepCure ONLY functions if the unit is in countdown mode. If the unit is in count up mode, StepCure will not function. (Exception is S2000 command set setting.)

For each step define the following:

- # – which UV LED Head is being controlled
- Time – the exposure time in seconds
- Level – the intensity of exposure, this will inherit the current status of the head connected. If the level is to be an irradiance, calibrate the head using an LS100, then set the irradiance value. If the level is to be a power, calibrate the head using an LS100P, then set the power value. If the stepcure program level differs from the connected head, an error will occur during stepcure execution.
- Dwell – the delay in seconds until the next step is performed
- Src – the source (input signal) which will trigger this step.

To clear any line, select the value in the # field and, using the arrow buttons, cycle down until it is blank.

*Example 1: Setting up a step profile*

	#	Time	Level	Dwell	Src
1	1	1.5	75	0.1	PLC1
2	1	5.0	100	2.0	PLC1
3	1	2.0	50	0.1	PLC1

A signal on the PLC pin "Enable 1" will turn on head 1 for 1.5s at 75%, then off for 0.1s, then 100% for 5.0s, then off for 2.0s, then 50% for 2.0s.

*Example 2: Controlling multiple heads with one input control signal*

	#	Time	Level	Dwell	Src
1	1	1.0	80	0.1	Fp2
2	2	1.5	65	0.1	Fp2
3	3	0.5	99	0.1	Fp2
4	4	10.0	25	0.1	Fp2

A signal on the foot pedal #2 connector will trigger an 80% exposure on head 1 for 1.0s, then a 65% exposure for 1.5s on head 2, then a 99% exposure on head 3 for 0.5s and then a 25% exposure on head 4 for 10.0s. There will be a 0.1s delay between each step.

*Example 3: Controlling multiple heads simultaneously with one input control signal*

## OmniCure® LX500 Series

### UV Curing Solutions

	#	Time	Level	Dwell	Src
1	1	1.0	80	0.1	Fp1
2	2	1.5	65	P	Fp1
3	3	0.5	99	P	Fp1

A signal on foot pedal #1 will run heads 1, 2 and 3 simultaneously.

Note: "P" will run the head at the same time as the above step

Note: If using the PLC Output, ensure that the head the PLC output is associated with is the first one listed in the parallel step. For the example above, the PLC output may be associated with head 1, but not with heads 2 or 3.

#### Example 4: Repeating a step cure program

	#	Time	Level	Dwell	Src
1	1	1.0	80	0.1	Fp1
2	2	1.5	65	P	Fp1
3	3	0.5	99	0.1	Fp1
4	R	Repeat 2 Times			Fp1

A signal on foot pedal #1 will run heads 1, 2 simultaneously, then head 3.

Note: "P" will run the head at the same time as the above step

Note: Pressing up button when changing Head #, after head 4 is R for repeat. A cycle may be repeated 2 to 10 times. Ensure the trigger source selected is correct for the program you wish to repeat.

#### Example 5: Using the S2000 command set

	#	Time	Level	Dwell	Src
1	1	--	--	--	PC
2	2	--	--	--	PC

Heads #1 and #2 will respond to the S2000 command set. When querying parameters, the values for head #1 will be returned only. Setting parameters will occur for both heads.

### Saving and Reusing Settings

Exposure settings are saved after running one exposure. StepCure settings are saved when exiting the StepCure screen.

Alternatively, to save the current exposure settings for future re-use, go to the SD Card Maintenance screen and press **Save**. A settings file will be created on the SD card with the name LX5SET.csv.

If a settings file already exists on the SD card, the LX500 detects it and, if the settings file is different than the present settings on the LX500, you will be asked if you want to load the saved settings on start up.

To manually load a settings file, scroll down to it in the file list on the SD Card Maintenance screen and press **Select**.

**Tip:** You can create a settings file on a PC and load onto an SD card. Refer section 8.7 SD Cards for the proper format for a settings file.

### 6.5 Running an Exposure from the Controller

1. Set the timer and/or intensity on the Main Control Panel, or use the StepCure screen to set up multiple step exposure parameters. Refer to [Section 6.4](#), "Defining Exposure Parameters".
2. Press **Start/Stop** on the front panel of the Controller to trigger an exposure.

All LEDs are in either count up or countdown mode:

- If the arrow in the status bar is pointing up, the unit is in count up mode. Each UV LED Head will be on until you press **Start/Stop** again or one of the other input control signals is toggled.
- If the arrow in the status bar is pointing down, the unit is in countdown mode. Each UV LED Head will be on for the duration of the set timer for that channel on the Main Control Panel screen.

### 6.6 Using the PC Interface with USB Connection

The LX500 can be commanded via a virtual RS-232 serial port. The serial port communication settings are not adhered to and instead communication occurs at the USB 2.0 speed of 12Mbps. All serial commands are available in an SDK which is available upon request.

The LX500 complies with the S2000 command set. Use the StepCure screen to select which LEDs will respond to this command set. To do this, the first row is the LED #, set the last row to PC. Refer to [Section 6.4](#), "Defining Exposure Parameters".

### 6.7 Interfacing with a PLC

#### PLC Input Signal Requirements

Signal	Max. Voltage	Max. Current
Enable 1-4 Front Panel Lock Count Up/Down Door Lock	5.0V	2mA (source)
Sync Out+ to Sync Out - Alarm+ to Alarm -	27V	100mA (sink)
Programmable Sync Out+ to Programmable Sync Out -	27V	500mA (sink)

#### PLC Connector Pin-Out




Figure 8 PLC Connector (looking at Controller)

Refer to chart below for the pinout on the LX500 system. The connector type used is DB25 Female.

Pin No.	Signal Name	Description
1	Enable 2	Shorting signal to GND will trigger channel 2 / 4.

## OmniCure® LX500 Series

UV Curing Solutions


Pin No.	Signal Name	Description
2	Enable 4	Leaving signal floating will not trigger channel 2 / 4. Logic for the signal can be inverted in the Settings screen. This signal is edge triggered. The function of this signal can be modified in the StepCure screen.
3	Front Panel Lock	Shorting signal to GND will lock out the front panel Leaving signal floating will leave the front panel unlocked Logic for the signal can be inverted in the Settings screen. This signal is edge triggered.
4	GND	
5	SYNCOUT <sub>1</sub> -	When the UV LED Head connected to channel 1-4 is on, pins are active. When the UV LED Head connected to channel 1-4 is off, pins are open. Logic for the signal can be inverted in the Settings screen.
6	SYNCOUT <sub>2</sub> -	
7	SYNCOUT <sub>3</sub> -	
8	SYNCOUT <sub>4</sub> -	
9	Programmable Sync Output -	When the UV LED head connected to channel 1 is on, pin is open. When the UV LED head connected to channel 1 is off, pin is open. A positive and negative delay can be programmed for these pins in the Settings screen.
10	ALARM (-)	When the LX500 has an error, pin is active. When the LX500 does not have an error, pin is open. Logic for the signal can be inverted in the Settings screen.
11	GND	
12	Reserved	
13	EMPTY	Not in use
14	Enable 1	Shorting signal to GND will trigger channel 1 / 3. Leaving signal floating will not trigger channel 1 / 3. Logic for the signal can be inverted in the Settings screen. This signal is edge triggered. The function of this signal can be modified in the StepCure screen.
15	Enable 3	
16	Count Up/Down	Shorting signal to GND will set system to "count up" mode Leaving signal floating will set system to "countdown" Logic for the signal can be inverted in the Settings screen. This signal is edge triggered
17	GND	
18	SYNCOUT <sub>1</sub> +	When the UV LED Head connected to channel 1-4 is on,

Pin No.	Signal Name	Description
19	SYNCOUT <sub>2</sub> +	pins are active. When the UV LED Head connected to channel 1-4 is off, pins are open. Logic for the signal can be inverted in the Settings screen.
20	SYNCOUT <sub>3</sub> +	
21	SYNCOUT <sub>4</sub> +	
22	Programmable Sync Output +	When the UV LED head connected to channel 1 is on, pin is active. When the UV LED head connected to channel 1 is off, pin is open. A positive and negative delay can be programmed for these pins in the Settings screen.
23	ALARM (+)	When the LX500 has an error, pin is active. When the LX500 does not have an error, pin is open. Logic for the signal can be inverted in the Settings screen.
24	GND	
25	Door Lock	Shorting signal to GND will enable all channels Leaving signal floating will disable all channels

**Table 1 PLC Connector Pin Out**

### Integrating Monitoring with External Circuitry

All inputs and outputs from the PLC interface are opto-isolated. The digital outputs are implemented using NPN logic. The equivalent circuit is shown in Figure 9 below.



**Figure 9 PLC digital output equivalent circuit**

### Running an Exposure

Optionally, use the StepCure screen (refer to [Section 6.4, "Defining Exposure Parameters"](#)) to override the default action of the PLC Enable 1-4 signal described below.

All LEDs are in either count up or countdown mode:

- If the arrow in the status bar is pointing up, the unit is in count up mode. The UV LED Head connected to channel 1-4 will be on for as long as the PLC Enable signal is held low or one of the other input control signals is toggled. Use count up mode when using an external timer for control of the LX500.
- If the arrow in the status bar is pointing down, the unit is in countdown mode. The UV LED Head connected to channel 1-4 will be on for the duration of the set timer for that channel on the Main Control Panel screen.

**Note:** Pulses must be longer than 10ms.

## 6.8 Using the Foot Pedal

The foot pedal does not disable the **Start/Stop** button and can be used simultaneously with the **Start/Stop** button on the front panel in any mode or setting. However, the **Start/Stop** button overrides all foot pedal configurations.



1. Connect the foot pedal to the port that corresponds with the UV LED Head on the back of the Controller (refer to [Figure 2](#)).
2. Use the StepCure screen (refer to [Section 6.4 "Defining Exposure Parameters"](#)) to set up the exposure parameters and, optionally, to override the default action described below.
3. Press the foot pedal to trigger an exposure.

All LEDs are in either count up or countdown mode:

- If the arrow in the status bar is pointing up, the unit is in count up mode. The UV LED Head connected to channel 1-4 will be on for as long as the foot pedal signal is held low or one of the other input control signals is toggled.
- If the arrow in the status bar is pointing down, the unit is in countdown mode. The UV LED Head connected to channel 1-4 will be on for the duration of the set timer for that channel on the Main Control Panel screen.

## 7 Troubleshooting

When a system error occurs an alarm notification will appear on the Controller display.

When head error occurs the error code will be displayed on the LED Head Information screen (refer to [Section 7.2, "Error Codes"](#)).

When an error occurs, the current active exposure will be terminated immediately.

On any head level error condition, the alarm PLC output will be activated. The alarm output will clear when all active alarms have been cleared, this may require a power reset to clear the error. If problems persist beyond these troubleshooting points, please contact Excelitas Technologies Service Department (refer to [Section 12, "Contact Information"](#)).

### 7.1 Symptoms and Possible Causes

*Service to be completed by qualified Excelitas Technologies personnel only!*

*Always wear optical safety glasses while troubleshooting. Never put hands or fingers in front of the lens. Serious personal injury may result.*

*Do not place any potentially combustible materials (e.g. paper, card, etc.) in front of the LED Head. Risk of fire and damage to the UV LED Head may result.*

#### **If the controller fails to power up or function properly:**

- Make sure the AC power cord is securely connected to a grounded (earthed mains socket) functional 3-pin outlet.
- Make sure the power supply cord is securely connected into the DC IN connector on the rear of the controller unit.

#### **If the UV LED head fails to turn on**

- Unsure that the Door Lock Jumper is securely connected to the PLC Connector. If you are providing your own PLC signals, ensure that Pins 24 and 25 are shorted together.
- Unsure the UV LED heads are securely connected to the back of the unit. Warning: turn off unit before plugging and unplugging the UV LED heads.
- Unsure that the StepCure screen is filled in properly or left blank.
- If you are using the start/stop button, ensure that the front panel has not been locked out from the PLC interface. If the front panel is locked out, a lock symbol will appear on the status bar.

#### **If the UV LED head gets a temperature fault**

- Ensure the ambient air temperature does not exceed 35C.
- If you are using the 55mm length heads, ensure that the clamp assembly is securely fitted.

#### **If the light intensity is too low:**

- Verify that the lens of the UV LED Head has been properly cleaned. Ensure Power is off.
- Verify that the intensity level of the UV LED Head is set to correct level.

### 7.2 Error Codes

Error	Description	Solution
<b>Bit 1 (1)</b>	The head(X) has exceeded the allowable operating temperature. (error code will remain present until cleared, however the head will be operative again when the temperature is below the maximum allowable). The channel select button is not required to clear this error. Subsequent triggers will clear the alarm if the temperature of the LED has recovered to below the over-temperature cut-off threshold. (This applies to the front panel Start/Stop button, Foot Pedal, and PLC I/O Start/Stop inputs).	<ul style="list-style-type: none"> <li>• Check position of clamp location on the head. Refer to Figure 5.</li> <li>• Check emission mode setting.</li> <li>• Check timing mode setting.</li> <li>• Check lens cleanliness.</li> </ul>
<b>Bit 2 (2)</b>	The head(X) has a short circuit.	<ul style="list-style-type: none"> <li>• Check LED head to controller connections.</li> </ul>
<b>Bit 3 (4)</b>	The head (X) has been disconnected after power on. (The head will remain disabled until the power is cycled).	<ul style="list-style-type: none"> <li>• Check LED head to controller connections.</li> </ul>
<b>Bit 4 (8)</b>	The head is open circuit. (The head will remain disabled until the power is cycled).	<ul style="list-style-type: none"> <li>• Check LED head to controller connections.</li> <li>• Check LED head wire for damage.</li> </ul>
<b>Bit 5 (16)</b>	The 1-wire memory chip for head (X) has an error. (The head will remain disabled until the power is cycled).	<ul style="list-style-type: none"> <li>• Check LED head to controller connections.</li> </ul>
<b>Bit 6 (32)</b>	No temperature reading from head (X). (The head will be deactivated until the power is cycled).	<ul style="list-style-type: none"> <li>• Check LED head to controller connections.</li> </ul>
<b>Bit 7 (64)</b>	Head is in a different controller serial # or different head connector position.	<ul style="list-style-type: none"> <li>• Power off unit and plug into controller serial # and head position specified by error message.</li> <li>• Press select to clear this error and accept the new connection.</li> </ul>
<b>Bit 8 (128)</b>	Intensity setting in the StepCure profile exceeds the capability of the head. Only occurs when the head is calibrated, and the StepCure setting is an irradiance or power setting. This also can occur if the setting in StepCure is an irradiance setting, and the head is calibrated in power, or vice versa.	<ul style="list-style-type: none"> <li>• Adjust the intensity setting in the StepCure profile.</li> <li>• Recalibrate the head, ensure that the peak intensity is found prior to calibration.</li> </ul>
<b>Err1</b>	Internal hardware error	<ul style="list-style-type: none"> <li>• Reset system.</li> <li>• If error code persists contact service center.</li> </ul>
<b>Err2</b>	Internal Communication failure	<ul style="list-style-type: none"> <li>• Reset system.</li> <li>• If error code persists contact service center.</li> </ul>
<b>Err3</b>	Non-Volatile Memory error	<ul style="list-style-type: none"> <li>• Reset system.</li> <li>• If error code persists contact service center.</li> </ul>
<b>Err4</b>	Not used	<ul style="list-style-type: none"> <li>• Not used</li> </ul>

## OmniCure® LX500 Series

UV Curing Solutions

---

Error	Description	Solution
<b>Err5</b>	Inadequate system performance	<ul style="list-style-type: none"><li>• The system cannot keep up with the control inputs, slow down the control inputs for proper operation.</li></ul>

**Table 2 Error Codes**

\* X = LED Input Ch1, Ch2, Ch3, Ch4

## 8 Care and Maintenance

**Note:** Excelitas Technologies recommends incorporating the cleaning of the LX500 system into the user/operator cleaning and maintenance schedule.

### 8.1 Cautions and Handling



**Caution:** Routine maintenance should only be completed by qualified personnel to avoid risk of injury/electrical shock to the end user. **No user serviceable components are located within the UV LED Head enclosure.**



**Warning:** Use UV radiation eye and skin protection during servicing if access to the UV LED modules is required during operation.

Risk Group 3 (IEC62471)

WARNING: UV emitted from this product. Avoid eye and skin exposure to unshielded product.

CAUTION: Possibly hazardous optical radiation emitted from this product. Do not stare into operating lamp.



**CAUTION UV:** Hazardous visible radiation is emitted from this product. Eye or skin irritation may result from exposure. Use appropriate shielding.

Ensure power supply is disconnected prior to making any changes to cabling configurations.

Operate the unit in a well-ventilated area with at least 5 inches (12cm) clearance on both sides of the Controller unit for proper air flow.

Avoid physical shocks or jarring of the LED Head, especially while the unit is operating.

The lens assembly includes a coated lens and lens holder. Improper handling and cleaning practices can damage surfaces or coatings. Note that damage to the surface of the glass or coating can degrade the Controller's performance. Follow proper handling and cleaning technique.

Always handle the head and lens assembly by the metal body; never touch the optical window assembly with your fingertips. The moisture or oils on your fingertips can damage the coating of the glass in the optical window assembly. Fingerprints left on an optical surface for an extended period of time can become a permanent stain. Even if you are wearing gloves, avoid touching the optical surface.

Never handle the glass lens/ window with metal implements or tweezers.

When not in use, store LED Head and Lens assemblies as described in [Section 5.4, "Storing LED Head and Lens Assemblies"](#). Place protective cap over the lens.

### 8.2 Cleaning Materials



**\*\*CAUTION\*\***

*Before using any solvent, consult the manufacturer's Materials Safety Data Sheets (MSDS) and your internal Health and Safety Advisor for proper handling and storage.*

- Pressurized gas (filtered dry nitrogen)
- Lint-free lens tissue, lint-free cotton swabs, lint or powder-free gloves or finger cots.
- An organic solvent, such as reagent-grade isopropyl alcohol, reagent-grade acetone, or lens cleaning solution.

### 8.3 Cleaning the LED Head and Lens Assembly



**\*\*WARNING\*\***

*Ensure power is OFF and the UV LED Head is cool before attempting any cleaning procedure.*

Regular cleaning of the optical surface is recommended for optimal performance.

Wear powder free latex gloves or finger cots suitable for the cleaning solution selected to protect hands from solvents and prevent window from contamination. Refer to the MSDS for guidance.

Blow off dirt and dust with pressurized air on front glass surface.

Use a lens tissue folded into quarters or a lint-free cotton swab. Saturate the lens tissue or swab with organic solvent as indicated in [Section 8.2, "Cleaning Materials"](#).

Using a new saturated tissue; clean LED Head by gently wiping the surface with a continuous stroke in one direction.

**Note:** Never apply a dirty tissue or swab to the UV LED Head.

**Note:** Never face the LED head upwards before the solvent has evaporated.

### 8.4 Cleaning the Controller



**\*\*CAUTION\*\***

*Do not use any alcohol- or acetone-based solutions to cleaning the controller housing.*

Ensure the controller is in the off position and the power supply is unplugged.

Use a mild soap solution as the cleaning media.

Dampen a clean lint free cloth, tissue or swab with the mild soap solution and wipe the exterior of the housing. Avoid excessive moisture when wiping in the air vent areas.

Using a new dampened tissue (or pre-saturated tissue), clean the optical window by gently wiping the surface with one continuous stroke in one direction.

**Note:** Never apply a dirty tissue or swab to the optical window of the controller.

Inspect the window to verify no residue is present.

### 8.5 Calibrating a UV LED Head



**\*\*CAUTION\*\***

*Do not stare directly at the light emitted from the LED array. This may be harmful, resulting in eye injury. Always use UV protective eyewear.*

The calibration process requires an Excelitas Technologies LS100 radiometer, or an LS100P radiometer. The LS100 is an irradiance only sensor that will calibrate the heads in W/cm<sup>2</sup>. The LS100P is a power only sensor that will calibrate the heads in W. Ensure the usage of the proper sensor for your application.

To calibrate the UV LED Heads connected to the LX500 controller:

1. Securely fixture the UV LED head you wish to calibrate.

2. Plug the LS100 or LS100P into the Radiometer connector on the front panel of the Controller (refer to [Figure 1](#)).
3. Ensure the Door Jumper connector is securely connected the PLC connector.
4. Position the HeadLS100 over the LS100/LS100P sensor.
5. For best results make sure the UV LED output is centeredcentred on the LS100 aperture.
6. On the Calibration screen, use the arrow keys to choose the appropriate Head number, then press Select. The Calibration Preparation Screen will appear. This screen allows for alignment of the head with the sensor, ability to clear the calibration, and cancelling out of the screen if only a radiometer was desired.
7. Press the start/stop button to turn on and off the head. Only the head selected for calibration will turn on and off. The head must be turned on in this screen before a calibration can occur, a reference reading is required for the calibration process. If the maximum intensity box is empty, the calibration will not be accessible.
8. The current LS100 reading in W/cm<sup>2</sup>, or LS100P reading in W, will appear along with the Maximum LS100/LS100P reading detected since this screen was entered. Use these values to ensure that the LS100 sensor is in the optimal position.
9. The up/down arrows can be used to increase or decrease the current set intensity of the head selected.
10. When properly aligned, pressed the Start Cal button to begin calibration.
11. The calibration can be cleared, also the screen can be exited with no modification to the current calibration. If either of these options are selected, the head will be turned off.

The progress of the calibration can be seen on the graph. Do not remove the LS100/LS100P while the UV LED head is being calibrated.

Once the calibration complete message appears, the calibration is complete. The intensity displayed on the "Main Control Panel" will be in W/cm<sup>2</sup> if calibrated with an LS100. If calibrated with an LS100P, the main control panel will show a decimal number and W for watts.

If a UV LED Head is moved to a different channel or unit, the calibration will be cleared. However the calibration will return if that head is plugged back into the original channel, as long as the head has not exceeded its calibration hours. Calibration hours will continue to decrement even when connected to a controller and/or head port it was not calibrated for.

The calibration information will be cleared after 100 hours.

The Calibration Preparation screen can also be used to clear the calibration of any head.

## 8.6 Viewing Log Files

The LX500 maintains two log files on a user-supplied micro SD card inserted on the front of the Controller (refer to [Figure 1](#)). If no SD card is inserted, no log file is generated or updated. All log files are automatically created in the root directory of the SD card, when an SD card is present.

- **Log File (LX5-LOG.CSV)** – This log file is updated each time the unit is turn off or on, settings are changed or a calibration has been performed. Each entry will have a time stamp. If there is no log file on an SD card, the LX5 will create one.
- **Calibration Log File (LX5CALLG.CSV)** – If an SD card is present during a calibration, a calibration log file will be created. All SD card files are in an Excel-compatible .csv format.

## 8.7 SD Cards

SD cards must be formatted as FAT12, FAT16 or FAT32 to be used in the LX500 system. The SD card system will save 2 different log files, setting files and firmware update files. Firmware update files have an EEC file

## OmniCure® LX500 Series

### UV Curing Solutions

extension; all other files are in a csv (Comma Separated Values) file format which is Microsoft Excel compatible.

#### Settings File

The settings file consists of two parts, The first row is the titles for the general settings, followed by the actual settings. The third row is the header for the StepCure section, followed by the StepCure rows. StepCure can have a maximum of 32 rows.

Item	Description
<b>General Settings</b>	
Date	The day the file was created, if created manually, enter today's date.
Time	The time the file was created, if created manually, enter present time.
LX500 SN	The serial number of the LX500 that the file was created on. If manually creating the file, enter 1.
Version	The version of the file, enter 0 for the current units.
Hd 1 Exp	The exposure time for head #1 in seconds.
Hd 1 Pwr	The intensity percentage of head #1.
Hd 1 Irr	The irradiance of head #1, only used if the head is calibrated, in W/cm2.
Hd 1 Cnt	The count up or down mode, enter 0 for down, 1 for up, all heads will be assigned the same value.
Hd 2 Exp	The exposure time for head #2 in seconds.
Hd 2 Pwr	The intensity percentage of head #2.
Hd 2 Irr	The irradiance of head #2, only used if the head is calibrated, in W/cm2.
Hd 2 Cnt	The count up or down mode, enter 0 for down, 1 for up, all heads will be assigned the same value.
Hd 3 Exp	The exposure time for head #3 in seconds.
Hd 3 Pwr	The intensity percentage of head #3.
Hd 3 Irr	The irradiance of head #3, only used if the head is calibrated, in W/cm2.
Hd 3 Cnt	The count up or down mode, enter 0 for down, 1 for up, all heads will be assigned the same value.
Hd 4 Exp	The exposure time for head #4 in seconds.
Hd 4 Pwr	The intensity percentage of head #4.
Hd 4 Irr	The irradiance of head #4, only used if the head is calibrated, in W/cm2.
Hd 4 Cnt	The count up or down mode, enter 0 for down, 1 for up, all heads will be assigned the same value.
PLC Invert	The logic for the PLC input, 0 is normal, 1 is inverted.
PLC Delay	The delay time for the PLC synch output, in ms, enter from -30000 to 30000.
Language	The language code for the system, currently only English is supported, enter 0 for the English language.
<b>StepCure Settings</b>	
Row #	The row # for the StepCure entry, not all rows must be filled, they can be skipped. Duplicate row number entries will result in the last entry being the only one used.
Hd #	The LED head # from 1 -> 4 for which the step is applicable to.
Exp Time	The exposure time in seconds.
Intensity	The intensity setting, in percentage.
Dwell	The dwell time for the step in seconds, enter 0 if the step is to run concurrently with the previous step.
Source	The source identifier,



Item	Description
	0 – None, violates step 1 – Foot Pedal 1 2 – Foot Pedal 2 3 – Foot Pedal 3 4 – Foot Pedal 4 5 – PLC Input 1 6 – PLC Input 2 7 – PLC Input 3 8 – PLC Input 4 9 – PC
Repeat	The number of times the program is to be repeated from 2>10.
Calibration	The calibration type the intensity value is for. If 0, the intensity value is % setting. If 1 the intensity value is irradiance in W/cm <sup>2</sup> . If 2, the intensity value is power in W.

### Log File

The log file is auto generated; the title row has the description for the columns as well as the following table:

Item	Description
Date	Date the log entry was added.
Time	Time the log entry was added.
Description	An English text description of the event.
Hd 1 Status	Error code for the head, see the error code description table for numbers to description.
Hd 1 Time	The exposure time setting in seconds.
Hd 1 Pwr	If a decimal number, the irradiance in W/cm <sup>2</sup> , if a whole number, the percentage setting.
Hd 1 Cnt	The exposure time count in seconds, that the head was in when the event occurred. E.g. if 5.0, it had 5.0 seconds remaining on the exposure.
Hd 1 Temp	The temperature of the head when the event occurred.
Hd 2 Status	Error code for the head, see the error code description table for numbers to description.
Hd 2 Time	The exposure time setting in seconds.
Hd 2 Pwr	If a decimal number, the irradiance in W/cm <sup>2</sup> , if a whole number, the percentage setting.
Hd 2 Cnt	The exposure time count in seconds that the head was in when the event occurred. E.g. if 5.0, it had 5.0 seconds remaining on the exposure.
Hd 2 Temp	The temperature of the head when the event occurred.
Hd 3 Status	Error code for the head, see the error code description table for numbers to description.
Hd 3 Time	The exposure time setting in seconds.
Hd 3 Pwr	If a decimal number, the irradiance in W/cm <sup>2</sup> , if a whole number, the percentage setting.
Hd 3 Cnt	The exposure time count in seconds that the head was in when the event occurred. E.g. if 5.0, it had 5.0 seconds remaining on the exposure.
Hd 3 Temp	The temperature of the head when the event occurred.
Hd 4 Status	Error code for the head, see the error code description table for numbers to description.
Hd 4 Time	The exposure time setting in seconds.
Hd 4 Pwr	If a decimal number, the irradiance in W/cm <sup>2</sup> , if a whole number, the percentage setting.
Hd 4 Cnt	The exposure time count in seconds, that the head was in when the event occurred. E.g.

## OmniCure® LX500 Series

UV Curing Solutions

---

Item	Description
	if 5.0, it had 5.0 seconds remaining on the exposure.
Hd 4 Temp	The temperature of the head when the event occurred.
System Status	The error code for the system, see the error code descriptions for number to description.

### Calibration Log File

The calibration log file keeps a record of all calibration performed on the system. Each time a calibration is performed a new record is added to the log file.

Item	Description
Date	Date the calibration log entry was added.
Time	Time the calibration log entry was added.
LX500 SN	The serial number of the LX500 that the calibration was performed on.
Hd #	The head number that was calibrated.
LS100 SN	The serial number of the LS100 sensor that was used for the calibration.
Cal Due	The calibration due date of the LS100 sensor that was used for the calibration.
Hd SN	The serial number of the LED head that was calibrated.
Wavelength	The wavelength of the LED head that was calibrated, also the wavelength the LS100 sensor was set to for the calibration.
Gain and %	The 10 gain settings and intensity percentages that were used for the calibration. The % vs irradiance reading is what generates the graph. The gain numbers are for Excelitas internal purposes only.
Tmp 1	The LED head temperature when the calibration was started, in °C
Tmp 2	The LED head temperature when the calibration completed, in °C
Cal Err	The error code from the calibration module, if 0, no errors occurred.
LS100 Err	The error code from the LS100 module, if 0, no errors occurred.
Status	Passed – Calibration completed successfully. Failed – Calibration failed, see Cal Err and LS100 Err for error codes.

## 9 Technical Specifications

### 9.1 Environmental Conditions

Operating Conditions	
Ambient Temperature	15°C to 35°C
Altitude	2000m max.
Atmospheric Pressure	700 to 1060 hPa
Relative Humidity:	15% to 85% (non-condensing)
Installation Category	II
Pollution Degree	2
Transport and Storage Conditions	
Temperature	-10 to +60°C
Relative Humidity	10% to 100% (non-condensing)
Atmospheric Pressure	500 to 1060 hPa

Table 3 Environmental Conditions

### 9.2 Controller Specifications

Specification		
Connectable Heads		1 to 4 individual LED heads.
Display		LCD TFT 2.4" Display
Mode Control		Mode and UP/Down button to control the display mode.
Timer		Programmable trigger, adjustable countdown range from 999.9s to 0.1, in 0.1s intervals or user selectable timing via trigger level.
Level		Adjustable output modes: 5%-100%
Start/Stop		Start/Stop button to control start or stop the LED Emission.
Radiometer Connector		Compatible with LS100 for output calibration
microSD Card		Save/load settings, save event or calibration log. (Supports Fat12, FAT16 and FAT32 file systems).
USB		USB control
Foot Pedal		Foot pedal to control start or stop the LED Emission.
External Control	Method	PLC
	External Input	Start/Stop, lock out front panel, count up/down
	External Output	Emissions, Alarm
Operating Voltage		Controller Supply Input: LX500-2: 12 VDC, 1.5A max LX500-4: 12 VDC, 3.0A max
AC Power Input		Input to AC adaptor: 100-240VAC , 50/60Hz, 0.21/0.085A for LX500-2 0.41/0.17A for LX500-4.
Ambient		Controller: 15°C to 35°C, 85% max. (non condensation).

Temperature/Humidity Range	
-------------------------------	--

Table 4: Controller Specifications

### 9.3 UV LED Head Specifications

Refer to OmniCure® LED Head Assembly Specification Guide 035-00638R for LED Head specifications.

The peak wavelength of the UV LED Head +/- 5nm.

The UV LED Head is capable of responding to PLC control signals in less than 1ms.

The UV LED Head stability during close loop feedback is  $\pm 3\%$

The UV LED Head calibration accuracy is  $\pm 5\%$  or  $\pm 0.25\text{W/cm}^2$  whichever is greater for irradiance and  $\pm 5\%$  or  $\pm 0.01\text{W}$  whichever is greater for power.


## 10 Regulatory Compliance

### 10.1 Product Safety and Electromagnetic Compatibility:

The LX5 Series has been tested and found to comply with product safety and electromagnetic compatibility requirements. For a complete list of tests and for certification details, please contact your OmniCure representative or visit

<http://www.excelitas.com/Pages/Product/OmniCure.aspx>

### CE Marking

Council Directive 2014/35/EU	Low Voltage Directive	
Council Directive 2014/30/EU	EMC Directive	
Council Directive 2012/19/EU	WEEE Directive	
Council Directive 2011/65/EU	RoHS	

### FCC Class A Digital Device or Peripheral – Information to User

#### Note:

This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at their own expense.

#### WARNING

Changes or modifications not expressly approved by Excelitas Technologies could void the user's authority to operate the equipment.

### 10.2 China RoHS



The symbol above indicates a product does not contain any restricted substances.

### 10.3 WEEE Directive (2012/19/EU)



The symbol above indicates that this product should not be disposed of along with municipal waste, that the product should be collected separately, and that a separate collection system exists for all products that contain this symbol within member states of the European Union.

The equipment that you bought has required the extraction and use of natural resources for its production. It may contain hazardous substances that could impact health and the environment.

In order to avoid the dissemination of those substances in our environment and to diminish the pressure on the natural resources, we encourage you to use the appropriate take-back systems. Those systems will reuse or recycle most of the materials of your end life equipment in a sound way.

The crossed-out wheeled bin symbol indicated above invites you to use those systems.

If you need more information on the collection, reuse and recycling systems, please contact your local or regional waste administration.

### 11 Warranty

Excelitas Technologies warrants the original purchaser for a period of one (1) full year, calculated from the date of purchase, that the equipment sold is free from defects in material and workmanship.

In the event of a claim under this warranty, the equipment is to be sent postage and carriage paid to the [Excelitas Technologies Service Centre](#). Returned equipment will not be received without a Return Authorization (RA) Number, issued by the appropriate Service Centre.

In order for us to serve you better, include a written description of the fault and the name and telephone number of a contact person who may be contacted for additional service related questions.

Any claims for units received with defects in material or workmanship must be reported to an authorized Excelitas Technologies Service Centre within 30 days from the original date of receipt and returned within 30 days of reporting to an authorized Excelitas Technologies Service Centre. Excelitas Technologies will repair or replace these reported defects free of charge. The equipment must be sent postage and carriage paid.

Package the equipment in its original shipping case or as appropriate to prevent damage during transport.

In the case of damage caused by wear and tear, careless handling, neglect, by the use of force or in the case of interventions and repairs not carried out by an Excelitas Technologies Authorized Service Centre, the warranty ceases to be valid. This warranty may not form the basis for any claims for damages, in particular not for compensation of consequential damages.

This warranty is not transferable.

**\*\*WARNING\*\***

*There are no field serviceable parts within the equipment. Opening the equipment will void the warranty.*

#### 11.1 Checking Your Warranty Status

The UV LED heads have a life expectancy of 10,000 hours. There are no alarms or warnings when a UV LED head exceeds 10,000 hours. You can check the operational hours for each head in the LED Head Information Screen.

#### 11.2 Replacement UV LED Module Warranty

If the LX500 system fails to power up during the warranty period of 10,000 operational hours, the UV LED Module will be replaced under warranty. In the event of a claim under this guarantee, the UV LED Head and/or LX500 Controller is to be sent postage and carriage paid, including a description of the fault, to the nearest appropriate [Excelitas Technologies Service Centre](#). Returned equipment will not be received without a Return Authorization (RA) Number, issued by the Service Centre. This warranty is non-transferable.

In the case of damage caused by careless handling, neglect, by the use of force or in the case of interventions and repairs not carried out by a Excelitas Technologies Service Centre to the LX500 system, the guarantee ceases to be valid.

#### 11.3 Returning your LX500 System to Excelitas Technologies for Service

Please make a note of the problem encountered, the steps followed to isolate the problem and the result of any trouble shooting steps taken.

Telephone the nearest Excelitas Technologies Service Centre to obtain a Return Authorization Number so that repairs may be completed quickly and efficiently. In North America, request for Return Authorization number can be made online at <http://www.excelitas.com/Pages/Support/Service-Instructions.aspx>.

## OmniCure® LX500 Series

UV Curing Solutions

---

Enclose details of the problem with the unit and return both to the Excelitas Technologies Service Centre. The unit should be returned in its original packaging if possible.

Include a phone number and contact person who may be reached for any additional service-related questions.



### 12 Contact Information

Excelitas Canada Inc.  
2260 Argentia Road  
Mississauga, Ontario  
L5N 6H7 CANADA

Tel.: +1 905 821-2600

Toll: +1 800 668-8752 (USA and Canada)

Fax: +1 905 821-2055

<http://www.excelitas.com/Pages/Product/OmniCure.aspx>

#### Technical Assistance:

[techsupport@excelitas.com](mailto:techsupport@excelitas.com)

<http://www.excelitas.com/Pages/Support/Service-Instructions.aspx>

For a complete listing of Authorized OmniCure Distributors and Service Centres please go to

<http://www.excelitas.com/Pages/Support/Service-Centers.aspx>.