

QUICK REFERENCE CARD

Aire Turn Bariatric Expandable

U-K-6XMS

1 Mattress Set-up Instructions


- Place mattress directly on the bed frame ensuring the hose is towards the footend.
- Tie the mattress straps to the loops located at the head of the bed, behind the backrest deck, to the attachment loops found at the central section of the patient deck, and to the openings found on each side of the liftable footend deck.
- **Note:** Ensure that there is no interference with the movement of the backrest, knee, bilateral extensions and foot of the bed.
- Hang the control unit on the footboard and attach the mattress connector to the controller.
- Ensure the CPR valve is closed.
- Connect the power cord to an electrical outlet or to the bed auxiliary outlet. The amber light will illuminate. The unit is in standby.

2 CPR Instructions

- Turn off the controller and turn the CPR valve to the open position.



3 Width & Length Extensions

For adequate bolster firmness, activate **MAX INFLATE**  mode between each width and length adjustment.

Width extension: 36", 42" or 48"

- Locate the valves at the foot of the mattress (patient right) and set valves to desired width according to visuals.



Length extension: 80", 84" or 88"
































- Locate the valves at the foot of the mattress (patient left) and set valves to desired length according to visuals.



This medical support surface is intended to be used by trained and qualified medical staff. Restrictions apply regarding its use in specific environments. Please refer to the user manual for detailed instructions, safety measures, and warnings.

For immediate assistance, please contact your local account executive.

3 Controller Instructions

-  Turns unit **ON/OFF** Green = On Amber = Standby
-  To activate **MAX INFLATE**, press  until max inflate  LED indicator illuminates. The compressor will come on and inflate the mattress. This function is also used to firm up the mattress to facilitate repositioning. It automatically deactivates after 20 minutes.
-  To do the **PATIENT SETUP**, press the  key and follow the on-screen instructions to set the patients height and weight. Patient comfort settings will be adjusted automatically.
-  To bypass automatic settings in the patient set up and to increase/decrease comfort pressure levels manually, press the **SOFT/FIRM**  keys.
-  The  keys on the control unit allow the caregiver to **INCREASE / DECREASE** therapy cycle time and rotation angle which is displayed on the control unit screen for the different therapies selected.
-  To program **ROTATION THERAPY**, press the ROTATION  key. Use the  keys to adjust the cycle time. To adjust the ROTATION ANGLE, press the ROTATION  key again and use the  keys to select the desired angle. Press the ROTATION  key once more to activate the rotation based on the selected settings.
-  To **PAUSE** rotation therapy cycle time, press the PAUSE  key.
-  To set **ALTERNATING PRESSURE THERAPY**, press the AP  key and use the  keys to adjust the cycle time between 1 and 99 minutes. Note: Alternating Pressure (AP) therapy is not available during Rotation Therapy.
-  Press select button  to access menu for other therapies available such as static, pulsation & massage. Follow the instruction on the screen to configure the desired settings.
-  To activate **LOW AIR LOSS**, press  until the LOW AIR LOSS  LED indicator illuminates.
-  To **SILENCE ALARM**, press the ALARM SILENCE KEY  until the LED indicator illuminates.
-  To set **LOCK OUT**, press and hold the LOCK KEY  until the LED indicator illuminates. To unlock, press and hold the LOCK KEY  until the LED indicator turns off.

