



CERTIFICATE OF INSTRUMENT CALIBRATION

INSTRUMENT: RSS-214DL, Heat Stress Monitor

CUSTOMER: ACKLANDS-GRAINER INC.

SERIAL NO: 1779

DATE OF CALIBRATION: AUGUST 9, 2000

IMAGING AND SENSING TECHNOLOGY HEREBY CERTIFIES THAT THE ABOVE IDENTIFIED INSTRUMENT WAS OPERATING WITHIN INSTRUCTION MANUAL ACCURACY LIMITS OVER THE TEMPERATURE RANGE GIVEN BELOW, ON THE DATE OF CALIBRATION, WITH ALL SENSORS ACTING AS DRY BULBS.

CALIBRATION DATA				
Nominal Test Temp. (°C)	Test Temp. (°C)	Wet Bulb (°C)	Dry Bulb (°C)	Globe (°C)
45.0	45.0	44.9	45.0	44.8

ATTEST: David L. Maythe Technician, 8/11/00
(Engineering) (Title) (Date)

ATTEST: [Signature] Engineer, QA, 8/11/00
(Quality) (Title) (Date)